



Special Accuracy (roll certificate required)
 @ 300°C +/- 0.40% (or +/- 1.0 °C)

Roll certificate required - Must be special limits

**VERY FAST RESPONSE
 BONDED JUNCTION
 TYPE K, J, T
 HAND HELD IMMERSION PROBE**

| | | |
|---|-----------------|---|
| TEMPERATURE CONTROLS PTY LTD | | QUALITY ENDORSED COMPANY Certificate No: QEC 14412 AS/NZS 9001:2008 |
| CONDITIONS | | SYDNEY: 7 YAMMA ST SEFTON, NSW 2162 PHONE: (02) 9721 8644 FAX : (02) 9738 9339 MELBOURNE: 8/280 Whitehall St YARRAVILLE, VIC 3013 PHONE: (03) 9687 0000 FAX : (03) 9687 1900 BRISBANE: 28A/125 Kerry Road, Archerfield Qld 4108 PHONE: 0418 222 868 EMAIL: sales@temperature.com.au |
| THIS DRAWING IS OUR EXCLUSIVE PROPERTY AND MUST NOT BE COPIED OR HANDED TO ANY OTHER PARTIES WITH OUT WRITTEN PERMISSION. IT SHALL BE RETURNED UPON REQUEST | | TITLE |
| SIZE : A4 | DATE : 14/05/14 | DWG NO : SD 069 STEPPED |
| SCALE : NTS | DRAWING BY R.B | REV : 02 |