



Special Accuracy (roll certificate required)  
 @ 300°C +/- 0.40% ( or +/- 1.0 °C)

Roll certificate required - Must be special limits

**Special Accuracy**  
**300°C +/- 0.40% ( or +/- 1.0 °C)**  
**TYRE TEMPERATURE THERMOCOUPLE**  
**TYPE K, J, T OR PT100**  
**HAND HELD IMMERSION PROBE**

**TEMPERATURE CONTROLS PTY LTD** Certificate No: QEC 14412  
 ISO 9001:2008

**CONDITIONS**

THIS DRAWING IS OUR EXCLUSIVE PROPERTY AND MUST NOT BE COPIED OR HANDED TO ANY OTHER PARTIES WITH OUT WRITTEN PERMISSION. IT SHALL BE RETURNED UPON REQUEST

SYDNEY : 7 YAMMA ST SEFTON , NSW 2162  
 PHONE : (02) 9721 8644 FAX : (02) 9738 9339  
 MELBOURNE : 8/280 Whitehall St  
 YARRAVILLE VIC 3013  
 PHONE : (03) 9687 0000 FAX : (03) 9687 1900  
 EMAIL : sales@temperature.com.au



TITLE

SIZE : A4	DATE : 07/10/10	<b>DWG NO: SD 085PVC</b>
SCALE : NTS	DRAWING BY R.B	<b>REV : 01</b>