



Special Accuracy
300°C +/- 0.40% (or +/- 1.0 °C)
TYPE K THERMOCOUPLE
COMPOST PROBE ASSEMBLY
FOR USE WITH TEMPCO305 DIGITAL INDICATOR

BONDED JUNCTION
 Special Accuracy (roll certificate required)
 @ 300°C +/- 0.40% (or +/- 1.0 °C)

Roll certificate required - Must be special limits

TEMPERATURE CONTROLS PTY LTD Certificate No: QEC 14412
 ISO 9001:2000

CONDITIONS

THIS DRAWING IS OUR
 EXCLUSIVE PROPERTY
 AND MUST NOT BE
 COPIED OR HANDED
 TO ANY OTHER PARTIES
 WITH OUT WRITTEN
 PERMISSION. IT SHALL
 BE RETURNED UPON
 REQUEST

SYDNEY : 7 YAMMA ST SEFTON , NSW 2162
 PHONE : (02) 9721 8644 FAX : (02) 9738 9339
 MELBOURNE : 8/280 Whitehall St
 YARRAVILLE VIC 3013
 PHONE : (03) 9687 0000 FAX : (03) 9687 1900
 EMAIL : sales@temperature.com.au



TITLE

SIZE : A4	DATE :18/03/03	DWG NO :SD 266
SCALE : NTS	DRAWING BY R.M	REV : 00