



Special Accuracy (roll certificate required)  
 @ 300°C +/- 0.40% ( or +/- 1.0 °C)

Roll certificate required - Must be special limits

Special Accuracy  
 300°C +/- 0.40% ( or +/- 1.0 °C)  
 TYPE K, J, T OR PT100  
 HAND HELD IMMERSION PROBE

**TEMPERATURE CONTROLS PTY LTD** Certificate No: QEC 14412  
 ISO 9001:2000

**CONDITIONS**

THIS DRAWING IS OUR EXCLUSIVE PROPERTY AND MUST NOT BE COPIED OR HANDED TO ANY OTHER PARTIES WITH OUT WRITTEN PERMISSION. IT SHALL BE RETURNED UPON REQUEST

SYDNEY : 7 YAMMA ST SEFTON , NSW 2162  
 PHONE : (02) 9721 8644 FAX : (02) 9738 9339  
 MELBOURNE : 8/280 Whitehall St  
 YARRAVILLE VIC 3013  
 PHONE : (03) 9687 0000 FAX : (03) 9687 1900  
 EMAIL : sales@temperature.com.au



TITLE

SIZE : A4	DATE :07/10/10	<b>DWG NO : SD 566</b>
SCALE : NTS	DRAWING BY R.B	<b>REV : 01</b>