

# Temperature Controls Pty Ltd



Quality  
Endorse  
Compan

7 YAMMA STREET SEFTON, NSW 2162 AUSTRALIA

Phone: 61 2 9721 8644

Fax: 61 2 9738 9339

Web: [www.temperature.com.au](http://www.temperature.com.au)

Email: [sales@temperature.com.au](mailto:sales@temperature.com.au)

ACN: 650 190 183 ABN: 966 501 901 83

ISO 9001

Lic: 14412

THERMOCOUPLES \* RTD SENSORS \* THERMOWELLS \* EXTENSION CABLES \* LEVEL SWITCHES  
1044 Internal pressure test

## **CERTIFICATE OF TEST**

**This certificate is to confirm the barstock screwed Thermowells**

Were subjected to the following test.

**Test Procedure:** Internal Hydraulic Test Procedure

Thermowell was subjected to a hydraulic pressure test 38,000 KPA for a duration of six minutes. The thermowell was individually tested on our Hydraulic pressure Comparator and compared with 150mm Ø DE WIT test pressure gauge serial No. 24702.  
Calibration certificate No. NAC/1973-E-3.

**Tag No:**

**Comment:** The Thermowell passed the test without failure.

**Date of Test:**

**Presented To:** **Order No:**

**Signed**

Michael Donnelly  
Testing Officer

MELBOURNE OFFICE: PHONE (03) 9687 0000 FAX: (03) 9687 1900  
PROUDLY MANUFACTURING IN AUSTRALIA