

TEMPERATURE CONTROLS PTY LTD

ACN: 650 190 183 ABN: 966 501 901 83

7 YAMMA STREET

SEFTON, NSW 2162 AUSTRALIA

PH: (02) 9721 8644 FAX: (02) 9738 9339 EMAIL: sales@temperature.com.au

www.temperature.com.au



HERMOCOUPLS * RTD SENSORS * THERMOWELLS * EXTENSION CABLES * LEVEL SWITCHES

Ref: 1056a

INSPECTION & TEST PLAN:

MIMS THERMOCOUPLE

CUSTOMER:

ORDER NO:

PRODUCT:

TAG NOS: As specified

WORKPLAN	REQUIREMENTS	SIGNATURE & RECORD
1. Document & record Material & Cable Size	Cable Size <input type="checkbox"/> Material <input type="checkbox"/>	
2. Document & Record Coil Number	Coil No: <input type="checkbox"/> Manufacturer: <input type="checkbox"/>	
3. Dimension Check & Verification	Confirm Diameter <input type="checkbox"/>	
4. Insulation Test & Measurement	Ambient 500 V DC > 100 M Ω <input type="checkbox"/> 300 C - 500 V - DC > 10 M Ω <input type="checkbox"/>	
5. Calibration Test & Measurement at 100 °C in Isotech Calibrator	Function Test & Calibration Test in Isotech Calibrator <input type="checkbox"/>	
6. Final Inspection	Q.A Manager to conduct Visual Dimensions Check and Audit Inspection & Test Plan <input type="checkbox"/>	
7. 3 rd Party Inspection	As required <input type="checkbox"/>	

Signed

.....
Michael Donnelly
Testing Officer